

Verification

Of Conformity with European Directives

Product: Vacuum Mixer

Type: AX-2000, AX-2000A

Applicant Manufacturer: Tianjin Aixin Medical Equipment Co., Ltd.

Address: 8A-701, No.15 Rongyuan Road, Huayuan New-Tech Industrial Park,

Tianjin, P.R. China, 300384

The submitted sample of the above product has been tested for CE

marking according to the following European Directive(s):

98/37/EC Machinery Directive 2006/95/EC Low Voltage Directive

2004/108/EC Electromagnetic Compatibility Directive

Standard(s) used for showing compliance with the essential requirements in the specified directive(s):

Standard(s): EN 60204-1:1997; EN ISO 12100-1:2003; EN ISO 12100-2:2003;

EN 1050: 1996; EN 55011: 2007+A2: 2007

Test report(s): 2008-1032

Issued by: Zhejiang Kezheng Electronic Information Product Testing Co., Ltd

Dates: 2008-1-24

The referred report(s) show that the product complies with standard(s) recognized as giving presumption of compliance with the essential requirements in the specified EU Directive(s). The CE marking as shown below may be affixed on the product.

Certificate No. No. 01211 Date: 22 Feb 2008

NQA

General manager

(Signature)

