



## Verification

Of Conformity with European Directives

Product: Dental Workstation  
Type: AX-JT3; AX-JT4; AX-JT6; AX-JT7; AX-YT1; AX-YT2;  
EM-LT1; EM-LT2; EM-LT3; EM-LT4; EM-LT5; EM-LT6  
Applicant Manufacturer: Tianjin Aixin Medical Equipment Co., Ltd.  
Address: 8A-701, No.15 Rongyuan Road, Huayuan New-Tech Industrial Park,  
Tianjin, P.R. China, 300384

**The submitted sample of the above product has been tested for CE marking according to the following European Directive(s):**

**98/37/EC Machinery Directive**  
**2006/95/EC Low Voltage Directive**

*Standard(s) used for showing compliance with the essential requirements in the specified directive(s):*

Standard(s): EN ISO 12100-1: 2003; EN ISO 12100-2: 2003;  
EN 60204-1: 1997; EN ISO 14121-1: 2007

Test report(s): 200808-J019

Issued: SNQA

Dates: 2008-08-21

The referred report(s) show that the product complies with standard(s) recognized as giving presumption of compliance with the essential requirements in the specified EU Directive(s).  
The CE marking as shown below may be affixed on the product.

Certificate No. **No. 01314**

Date: **29 Oct 2008**

NQA

General manager (Signature)

